

A Szegedi Tudományegyetem [redacted] /2012/SZTE. számú,
Biológiai Erősítő készülék beszerzése tárgyú közbeszerzési eljárásának műszaki
 minimumfeltételei

A táblázat kitöltése, azaz a megajánlott eszköz műszaki paramétereinek alábbi formában történő megadása kötelező. Az „*azonos / eltérő*” oszlopban, az adott sorban meghatározott paraméternek való pontos megfelelés esetén kell az „*azonos*” megjelölést feltüntetni, bármilyen irányú eltérés esetén pedig az „*eltérő*” jelölést, mindkét esetben rögzítve a következő oszlopban a konkrét paramétereket.

Az eszközzel szemben támasztott minimum követelmények	Ajánlat megfelelése a minimum követelményeknek	
	azonos/eltérő	Az ajánlat szerinti eszköz műszaki paramétere
I. Myocam Contractility Subsystem		
Microscope Camera System Including High-Speed Contractility Camera with Controller, Image Splitter Boks with 585 nm Dichroic Filter and a Photon Counting Photomultiplier Tube		
II. DigiData 1440 ADC Unit		
DIGIDATA 1440 LOW-NOISE DATA ACQUISITION SYSTEM, 16-BIT, 250 KSPS, USB 2.0 PCI card, USB 2.0 cable, External auto-switching power supply, Power cable, AxoScope 10 software, Manual		
III. pClamp 10 Upgrade from pClamp 8		
pClamp 10 CNS software upgrade from pClamp 8 with Minidigi.		